Ordering Information:

optix

Email orders to: sales@xsoptix.com

Fax orders to: 800-878-7282

800 Village Walk #316 Guilford, CT 06437 Ph: 203-401-8093



Revision 0.81

## SINGLE FREQUENCY LASER DIODES **Stabilized Ridge Waveguide Laser**



2021-02-19

#### General Product Information

Product	Application
780 nm Wavelength Stabilized Laser	Metrology
with hermetic 8-Pin TO Package (RoHS compliant)	Interferometry
including Monitor Diode, Thermoelectric Cooler and Thermistor	



Parameter	Symbol	Unit	min	typ	max
Storage Temperature	Ts	°C	-40		85
Operational Temperature at Case	T <sub>C</sub>	°C	-20		75
Operational Temperature at Laser Chip	T <sub>LD</sub>	°C	0		50
Forward Current	١ <sub>F</sub>	mA			200
Reverse Voltage	V <sub>R</sub>	V			2
Output Power	P <sub>opt</sub>	mW			100
TEC Current	I <sub>TEC</sub>	А			1.0
TEC Voltage	V <sub>TEC</sub>	V			1.0

### Recommended Operational Conditions

Parameter	Symbol	Unit	min	typ	max
Operational Temperature at Case	T <sub>case</sub>	°C	-20		65
Operational Temperature at Laser Chip	T <sub>LD</sub>	°C	15		45
Forward Current	I <sub>F</sub>	mA			180
Output Power	P <sub>opt</sub>	mW	20		80

#### Characteristics at T<sub>LD</sub> = 25° C at BOL

Parameter	Symbol	Unit	min	typ	max
Center Wavelength	$\lambda_{C}$	nm	778	780	783
Selectable Linewidth	Δλ	pm			0.1
Overall Linewidth	Δλ	nm			0.2
Temperature Coefficient of Wavelength	dλ / dT	nm / K		0.06	
Current Coefficient of Wavelength	dλ / dI	nm / mA		0.003	
Sidemode Supression Ratio	SMSR	dB	30	45	

eagleyard Photonics GmbH

12489 Berlin GERMANY Rudower Chaussee 29 www.toptica-eagleyard.com info@toptica-eagleyard.com fon +49.30.6392 4520



#### Measurement Conditions / Comments

Stress in excess of one of the Absolute Maximum Ratings may damage the laser. Please note that a damaging optical power level may occur although the maximum current is not reached. These are stress ratings only, and functional operation at these or any other conditions beyond those indicated under Recommended Operational Conditions is not implied.

# Measurement Conditions / Comments measured by integrated Thermistor Measurement Conditions / Comments single mode operation (see p. 4) multi mode operation (see p. 4)

under single mode condition

#### This data sheet is subject to change without notice. © eagleyard Photonics

Revision 0.81

## SINGLE FREQUENCY LASER DIODES Stabilized Ridge Waveguide Laser

Characteristics at $T_{LD} = 25^{\circ}$ (	C at BOL				cont'd
Parameter	Symbol	Unit	min	typ	max
Laser Current @ P <sub>opt</sub> = 80 mW	I <sub>LD</sub>	mA			180
Slope Efficiency	η	W / A	0.6	1.0	1.4
Threshold Current	I <sub>th</sub>	mA			70
Divergence parallel (FWHM)	$\Theta_{  }$	0		8	
Divergence perpendicular (FWHM)	$\Theta_{\perp}$	0		21	

Measurement Conditions / Comments
parallel to Pin 1 - Pin 6 plane (see p. 3)
perpendicular to Pin 1 - Pin 6 plane (see p. 3)

Measurement Conditions / Comments

 $U_R = 5 V$ 

#### Monitor Diode

Symbol	Unit	min	typ	max
I <sub>mon</sub> / P <sub>opt</sub>	µA/mW		2	
	- <b>,</b>	Symbol Unit I <sub>mon</sub> / P <sub>opt</sub> µA/mW	,	3 31

#### Thermoelectric Cooler

Parameter	Symbol	Unit	min	typ	max
Current	I <sub>TEC</sub>	А		0.4	
Voltage	U <sub>TEC</sub>	V		0.4	
Power Dissipation (total loss at case)	Ploss	W		0.4	
Temperature Difference	ΔΤ	К			40

## Thermistor (Standard NTC Type)

ResistanceR $k\Omega$ 10Beta Coefficient $\beta$ 3930Steinhart & Hart Coefficient AA1.029 x 10 $^{-3}$ Steinhart & Hart Coefficient BB2.510 x 10 $^{-4}$	Parameter	Symbol	Unit min typ			
Steinhart & Hart Coefficient A A 1.029 x 10 <sup>-3</sup>	Resistance	R	kΩ		10	
	Beta Coefficient	β			3930	
Steinhart & Hart Coefficient B B 2 510 x 10 <sup>-4</sup>	Steinhart & Hart Coefficient A	А			1.029 x 10	-3
	Steinhart & Hart Coefficient B	В			2.510 x 10	-4
Steinhart & Hart Coefficient C C 1.051 x 10 <sup>-7</sup>	Steinhart & Hart Coefficient C	С			1.051 x 10	-7

Measurement Conditions / Comments
$P_{opt} = 80 \text{ mW}, \Delta T = 20 \text{ K}$
$P_{opt} = 80 \text{ mW}, \Delta T = 20 \text{ K}$
$P_{opt} = 80 \text{ mW}, \Delta T = 20 \text{ K}$
$P_{opt} = 80 \text{ mW}, \Delta T =  Tcase - TLD $

$T_{LD} = 25^{\circ} C$	
$R_1/R_2 = e^{\beta(1/T_1\cdot1/T_2)}$ at $T_{LD} =$	0° 50° C
$1/T = A + B(\ln R) + C(\ln R)^{3}$	
T: temperature in Kelvin	
R: resistance at T in Ohm	

Rudower Chaussee 29 12489 Berlin GERMANY www.toptica-eagleyard.com info@toptica-eagleyard.com fon +49.30.6392 4520

This data sheet is subject to change without notice. © eagleyard Photonics








2021-02-19

Revision 0.81

## SINGLE FREQUENCY LASER DIODES Stabilized Ridge Waveguide Laser


## Pin Assignment

1	Laser Diode Anode	5	Thermistor
2	Laser Diode Cathode	6	Thermistor
3	Thermoelectric Cooler (-)	7	Photo Diode Anode
4	Thermoelectric Cooler (+)	8	Photo Diode Cathode



All 8 pins are isolated from case.

### Package Drawings



eagleyard Photonics GmbH Rudower Chaussee 29 12489 Berlin GERMANY www.toptica-eagleyard.com info@toptica-eagleyard.com fon +49.30.6392 4520


This data sheet is subject to change without notice. © eagleyard Photonics

Revision 0.81

## SINGLE FREQUENCY LASER DIODES Stabilized Ridge Waveguide Laser

#### Typical Measurement Results

#### Spectral maps at 25° C and 35° C



The spectral maps show the power spectral density at different operating modes. The graphs illustrate that the laser exhibits single and multi mode behavior under different operational conditions. The spectral maps may differ from part to part. Single mode operation can be achieved by selecting the appropriate laser current and temperature.


#### Unpacking, Installation and Laser Safety

Unpacking the laser diodes should only be done at electrostatic safe workstations (EPA). Though protection against electro static discharge (ESD) is implemented in the laser package, charges may occur at surfaces. Please store this product in its original package at a dry, clean place until final use. During device installation, ESD protection has to be maintained.

The RWS laser is sensitive against optical feedback, so an optical isolator may be required in order to avoid any disturbance of the emission spectrum. Operating at moderate temperatures on proper heat sinks will contribute to a long lifetime of the diode.

Avoid direct and/or indirect exposure to the free running beam. Collimating and focussing the free running beam with optics as common in optical instruments will increase threat to the human eye.

Performance figures, data and any illustrative material provided in this specification are typical and must be specifically confirmed in writing by eagleyard Photonics before they become applicable to any particular order or contract. In accordance with the eagleyard Photonics policy of continuous improvement specifications may change without notice.



Complies with 21 CFR 1040.10 and 1040.40

eagleyard Photonics GmbH Rudower Chaussee 29 12489 Berlin GERMANY www.toptica-eagleyard.com info@toptica-eagleyard.com fon +49.30.6392 4520

This data sheet is subject to change without notice. © eagleyard Photonics



TOPTICA