



# EYP-TBR-1064-02000-6000-BTW09-0000

Revision 0.70



2023-01-23


## SINGLE FREQUENCY LASER DIODES Distributed Bragg Reflector Laser with integrated Amplifier



### General Product Information

Product	Application
1064 nm DBR Laser	Raman Spectroscopy
with monolithically integrated Tapered Amplifier (TA)	Metrology
hermetic 14 Pin Butterfly Housing (RoHS compliant)	Nd:YAG Replacement
including Thermoelectric Cooler and Thermistor	EDFA Pumping

**μMOPA**



### Absolute Maximum Ratings

Parameter	Symbol	Unit	min	typ	max
Storage Temperature	$T_S$	°C	-40		85
Operational Temperature at Case	$T_C$	°C	-20		75
Operational Temperature at Laser Chip	$T_{LD}$	°C	10		50
Forward Current DBR	$I_{DBR}$	mA			500
Forward Current TA	$I_{TA}$	A			7.5
Reverse Voltage DBR	$V_{R,DBR}$	V			2
Reverse Voltage TA	$V_{R,TA}$	V			2
Output Power	$P_{opt}$	W			2.5
TEC Current	$I_{TEC}$	A			2.5
TEC Voltage	$V_{TEC}$	V			5.0

#### Measurement Conditions / Comments

Stress in excess of one of the Absolute Maximum Ratings may damage the laser. Please note that a damaging optical power level may occur although the maximum current is not reached. These are stress ratings only, and functional operation at these or any other conditions beyond those indicated under Recommended Operational Conditions is not implied.

### Recommended Operational Conditions

Parameter	Symbol	Unit	min	typ	max
Operational Temperature at Case	$T_C$	°C	0		50
Operational Temperature at Laser Chip	$T_{LD}$	°C	15	25	35
Forward Current DBR	$I_{DBR}$	mA			450
Forward Current TA	$I_{TA}$	A			7.0
Output Power	$P_{opt}$	W			2.0

#### Measurement Conditions / Comments

### Characteristics

Parameter	Symbol	Unit	min	typ	max
Center Wavelength	$\lambda_C$	nm	1063	1064	1065
Spectral Width (FWHM)	$\Delta\lambda$	pm		3	
Sidemode Suppression Ratio	SMSR	dB	30		

#### Measurement Conditions / Comments

see images on page 4  
apart from mode-hops (see Spectral Map on page 4)  
 $P_{opt} = 2 \text{ W}$

## SINGLE FREQUENCY LASER DIODES

### Distributed Bragg Reflector Laser with integrated Amplifier



#### Characteristics

cont'd

Parameter	Symbol	Unit	min	typ	max
Temperature Coefficient of Wavelength	$d\lambda / dT$	nm / K		0.08	
Current Coefficient of Wavelength	$d\lambda / dI_{DBR}$	nm / mA		0.001	
Current Coefficient of Wavelength	$d\lambda / dI_{TA}$	nm / A		0.035	
Laser Current @ $P_{opt} = 2.0$ W	$I_{TA}$	A			7.0
Slope Efficiency	$\eta$	W / A		0.8	
Threshold Current	$I_{th, TA}$	A		3	
Divergence parallel (FWHM)	$\Theta_{  }$	mrad		2	
Divergence perpendicular (FWHM)	$\Theta_{\perp}$	mrad		2	
Beam Diameter horizontal ( $1/e^2$ )	$d_{  }$	mm		1	
Beam Diameter vertical ( $1/e^2$ )	$d_{\perp}$	mm		1	
Degree of Polarization	DOP	%		90	

#### Measurement Conditions / Comments

Laser Forward Current DBR

Laser Forward Current TA

Amplifier

parallel to the base plate of the housing (see p. 3)

perpendicular to base plate of the housing (see p. 3)

parallel to the base plate of the housing (see p. 3)

perpendicular to base plate of the housing (see p. 3)

 $P_{opt} = 2$  W; E field parallel to the base plate

#### Thermoelectric Cooler

Parameter	Symbol	Unit	min	typ	max
Current	$I_{TEC}$	A			2.5
Voltage	$U_{TEC}$	V			5.0
Power Dissipation (total loss at case)	$P_{loss}$	W		10	
Temperature Difference	$\Delta T$	K			25

#### Measurement Conditions / Comments

 $P_{opt} = 2$  W,  $\Delta T = 20$  K $P_{opt} = 2$  W,  $\Delta T = 20$  K $P_{opt} = 2$  W,  $\Delta T = 20$  K $P_{opt} = 2$  W,  $\Delta T = |T_{case} - T_{LD}|$ 

#### Thermistor (Standard NTC Type)

Parameter	Symbol	Unit	min	typ	max
Resistance	R	k $\Omega$		10	
Beta Coefficient	$\beta$			3892	
Steinhart & Hart Coefficient A	A			$1.1293 \times 10^{-3}$	
Steinhart & Hart Coefficient B	B			$2.3410 \times 10^{-4}$	
Steinhart & Hart Coefficient C	C			$8.7755 \times 10^{-8}$	

#### Measurement Conditions / Comments

 $T_{LD} = 25^{\circ}$  C $R_1 / R_2 = e^{\beta(1/T_1 - 1/T_2)}$  at  $T_{LD} = 0^{\circ} \dots 50^{\circ}$  C $1/T = A + B(\ln R) + C(\ln R)^3$ 

T: temperature in Kelvin

R: resistance at T in Ohm

# EYP-TBR-1064-02000-6000-BTW09-0000

Revision 0.70




2023-01-23

## SINGLE FREQUENCY LASER DIODES Distributed Bragg Reflector Laser with integrated Amplifier






### Pin Assignment

1	Thermoelectric Cooler (+)	14	Thermoelectric Cooler (-)
2	Thermistor +	13	not connected
3	not connected	12	Amplifier Cathode
4	(Thermistor +)	11	Amplifier Cathode
5	Thermistor -	10	Amplifier Anode
6	DBR Laser Cathode	9	Amplifier Anode
7	DBR Laser Anode	8	not connected



### Package Drawings



Caution. Excessive mechanical stress on the package can lead to a damage of the laser.

See [instruction manual](#) on [www.eagleyard.com](http://www.eagleyard.com)


SWZ-23-0116-1755

## SINGLE FREQUENCY LASER DIODES Distributed Bragg Reflector Laser with integrated Amplifier



### Typical Measurement Results

Spectral Map



Performance figures, data and any illustrative material provided in this specification are typical and must be specifically confirmed in writing by eagleyard Photonics before they become applicable to any particular order or contract. In accordance with the eagleyard Photonics policy of continuous improvement specifications may change without notice.


### Unpacking, Installation and Laser Safety

Unpacking the laser diodes should only be done at electrostatic safe workstations (EPA). Though protection against electro static discharge (ESD) is implemented in the laser package, charges may occur at surfaces. Please store this product in its original package at a dry, clean place until final use. During device installation, ESD protection has to be maintained.

The TBR laser is sensitive against optical feedback, so an optical isolator may be required in order to avoid any disturbance of the emission spectrum. Operating at moderate temperatures on proper heat sinks will contribute to a long lifetime of the diode.

Avoid direct and/or indirect exposure to the free running beam. Collimating and focussing the free running beam with optics as common in optical instruments will increase threat to the human eye.

Performance figures, data and any illustrative material provided in this specification are typical and must be specifically confirmed in writing by eagleyard Photonics before they become applicable to any particular order or contract. In accordance with the eagleyard Photonics policy of continuous improvement specifications may change without notice.



IEC-60825-0

